



University of Ruhuna
Wellamadama
Matara

30.06.2022

The Manager,
.....


Invitation for Bids (IFB)

Contract: Supply, Delivery, Installation, Commissioning and Testing of Laboratory Equipment for Faculty of Medicine, University of Ruhuna.

IFB No: AHEAD/RUH/STEM3/GOODS/2022/NS-182 II

The Chairman, Department Procurement Committee on behalf of the University of Ruhuna invites sealed bid from you for **Supply, Delivery, Installation, Commissioning and Testing of Laboratory Equipment** indicated in the annexed bid document.

01. Bidding will be conducted through National Shopping method of procurement.
02. You may obtain further information if any from the **Deputy Director- Procurement (AHEAD Project) Supplies Branch, University of Ruhuna, Wellamadama, Matara. (T.P 041- 2227027/ Fax No. 041 – 2227027)**
03. Bids must be delivered by mail in sealed envelopes addressed to **Director/OTS, AHEAD Project-OTS Office, University of Ruhuna, Wellamadama, Matara before 02.00 p.m. on July 15, 2022.** Late bids will be rejected.
04. This contract is subject to,
 - I. Providing of performance security of 10% of the total contract price in accordance with the format attached and issued with this in favor of the **Vice Chancellor, University of Ruhuna.**
 - II. Entering in to the formal agreement with the University of Ruhuna.

 30/06/2022
.....
Director – Operational Technical Secretariat,
AHEAD Project,
University of Ruhuna,
Wellamadama,
Matara.

Dr. V. L. KRISHNAN
Deputy Director - Procurement
(Operational Technical Secretariat)
AHEAD Project
University of Ruhuna
Wellamadama, Matara.



UNIVERSITY OF RUHUNA

**PROCUREMENT OF GOODS UNDER
NATIONAL SHOPPING PROCEDURES**

**INVITATION FOR QUOTATIONS
FOR**

**Supply, Delivery, Installation, Commissioning and Testing of
Laboratory Equipment for Faculty of Medicine, University of Ruhuna.**

IFB No: AHEAD/RUH/STEM3/GOODS/2022/NS-182 II

*Mr. G. V. L. Krishnan
Deputy Director - Procurement
Operations Technical Services
AHEAD Project
University Of Ruhuna
Wellamadana, Matara.*

Section I. Instructions to Vendors (ITV)

A: General	
1. Scope of Bid	1.1 The Purchaser named in the Data Sheet invites you to submit a quotation for the supply of Goods as specified in Section III Schedule of Requirements.
B: Contents of Documents	
2. Contents of Documents	2.1 The documents consist of the Sections indicated below. <ul style="list-style-type: none"> • Section I. Instructions to Vendors (ITV) • Section II. Data Sheet • Section III. Schedule of Requirements • Section IV. Technical Specifications & Compliance with Specifications with price schedule • Section V. Quotation submission Form(s)
C: Preparation of Quotation	
3. Documents Comprising your Quotation	3.1 The Quotation shall comprise the following: <ul style="list-style-type: none"> (a) Quotation Submission Form and the Price Schedules; (b) Technical Specifications & Compliance with Specifications
4. Quotation Submission Form and Price Schedules	4.1 The vendor shall submit the Quotation Submission Form using the form furnished in Section V. This form must be completed without any alterations to its format, and no substitutes shall be accepted. All blank spaces shall be filled in with the information requested. 4.2 Alternative offers shall not be considered. The vendors are advised not to quote different options for the same item but furnish the most competitive among the options available to the bidder.
5. Prices and Discounts	5.1 Unless specifically stated in Data Sheet, all items must be priced separately in the Price Schedules. 5.2 The price to be quoted in the Quotation Submission Form shall be the total price of the Quotation, including any discounts offered.

Mrs. D. V. L. Krishnam
 Deputy Director - Procurement
 Operations Technical Secretariat
 ANEAD Project
 University Of Ruhuna
 Welisara, Sri Lanka

	<p>5.3 The applicable VAT shall be indicated separately.</p> <p>5.4 Prices quoted by the vendor shall be fixed during the vendor's performance of the Contract and not subject to variation on any account. A Quotation submitted with adjustable price shall be treated as nonresponsive and may be rejected.</p>
6. Currency	6.1 The vendors shall quote only in Sri Lanka Rupees.
7. Documents to Establish the Conformity of the Goods	<p>7.1 The vendor shall furnish as part of its quotation the documentary evidence that the Goods conform to the technical specifications and standards specified in Section IV, "Technical Specifications & Compliance with Specifications".</p> <p>7.2 The documentary evidence may be in the form of literature, drawings or data, and shall consist of a detailed item by item description of the essential technical and performance characteristics of the Goods, demonstrating substantial responsiveness of the Goods to the technical specifications, and if applicable, a statement of deviations and exceptions to the provisions of the Technical Specifications given.</p> <p>7.3 If stated in the Data Sheet the vendor shall submit a certificate from the manufacturer to demonstrate that it has been duly authorized by the manufacturer or producer of the Goods to supply these Goods in Sri Lanka.</p>
8. Period of Validity of quotation	8.1 Quotations shall remain valid for the period of sixty (60) days after the quotation submission deadline date.
9. Format and Signing of Quotation	9.1 The quotation shall be typed or written in indelible ink and shall be signed by a person duly authorized to sign on behalf of the vendor.
D: Submission and Opening of Quotation	
10. Submission of Quotation	<p>10.1 Vendors may submit their quotations by Registered post in sealed envelopes addressed to the Purchaser and bear the specific identification of the quotation number.</p> <p>10.2 If the quotation is not sealed and marked as required, the Purchaser will assume no responsibility for the misplacement or premature opening of the quotation.</p>
11. Deadline for Submission of Quotation	11.1 Quotation must be received by the Purchaser at the address set out in Section II, "Data Sheet", and not later than the date and time as specified in the Data Sheet.
12. Late Quotations	12.1 The Purchaser shall reject any quotation that arrives after the deadline for submission of quotations, in accordance with ITV Clause 11.1 above.

13. Opening of Quotations	<p>13.1 The Purchaser shall conduct the opening of quotation in public at the address, date and time specified in the Data Sheet.</p> <p>13.2 A representative of the bidders may be present and mark his attendance.</p>
E: Evaluation and Comparison of Quotations	
14. Clarifications	<p>14.1 To assist in the examination, evaluation and comparison of the quotations, the Purchaser may, at its discretion, ask any vendor for a clarification of his quotation. Any clarification submitted by a vendor in respect to his quotation which is not in response to a request by the Purchaser shall not be considered.</p> <p>14.2 The Purchaser's request for clarification and the response shall be in writing.</p>
15. Responsiveness of Quotations	<p>15.1 The Purchaser will determine the responsiveness of the quotation to the documents based on the contents of the quotation received.</p> <p>15.2 If a quotation is evaluated as not substantially responsive to the documents issued, it may be rejected by the Purchaser.</p>
16. Evaluation of quotations	<p>16.1 The Purchaser shall evaluate each quotation that has been determined, to be substantially responsive.</p> <p>16.2 To evaluate a quotation, the Purchaser may consider the following:</p> <ul style="list-style-type: none"> (a) the Price as quoted; (b) price adjustment for correction of arithmetical errors; (c) price adjustment due to discounts offered. <p>16.3 The Purchaser's evaluation of a quotation may require the consideration of other factors, in addition to the Price quoted if stated in Section II, Data Sheet. These factors may be related to the characteristics, performance, and terms and conditions of purchase of the Goods.</p>
17. Purchaser's Right to Accept any Quotation, and to Reject any or all Quotations	<p>17.1 The Purchaser reserves the right to accept or reject any quotation, and to annul the process and reject all quotations at any time prior to acceptance, without thereby incurring any liability to bidders.</p>

Mr. D. V. L. Krishanu
 Joint Director - Procurement
 Operations Technical Secretariat
 AHEAD Project
 University Of Ruhuna
 Wellamadulla, Sri Lanka

F: Award of Contract	
18. Acceptance of the Quotation	18.1 The Purchaser will accept the quotation of the vendor whose offer has been determined to be the lowest evaluated bid and is substantially responsive to the documents issued.
19. Notification of acceptance	19.1 Prior to the expiration of the period of validity of quotation, the Purchaser will notify the successful vendor, in writing, that his quotation has been accepted.
20. Signing of Contract	20.1 Within (14) days after notification, the Purchaser shall complete the Agreement, and in form the successful Bidder have to sign it. 20.2 Within fourteen (14) days of receipt of such information, the successful Bidder shall sign the Agreement.
21. Performance Security.	21.1 Within fourteen (14) days of the receipt of notification of award from the Purchaser, the successful Bidder is required furnish the Performance Security in accordance with the CC, using for that purpose the Performance Security Form included in Section VII contract forms, The Employer shall promptly notify the name of the winning Bidder to each unsuccessful Bidder and discharge the Bid Security of the unsuccessful bidders. 21.2 Failure of the successful Bidder to submit the above mentioned Performance Security or sign the contract shall constitute sufficient grounds for the annulment of the award and forfeiture of the Bid Security or execution of the Bid-Securing Declaration. In that event the Purchaser may award the Contract to the next lowest evaluated Bidder, whose offer is substantially responsive and is determined by the purchaser to be qualified to perform the contract satisfactorily.

Ms. D. V. L. Krishani
 Deputy Director - Procurement
 Operations Technical Secretariat
 AHEAD Project
 University Of Ruhuna
 Galle, Sri Lanka.

Section II: Data Sheet

ITV Clause Reference	
1.1	The Purchaser is: Vice Chancellor Address : University of Ruhuna, Wellamadama Matara.
5.1	If the bidder is allowed to quote for lesser number of items than the total number of items specified, indicate the details.
7.3	Manufacturer's Authorization is required.
11.1	Address for submission of quotations is Director/ OTS, AHEAD Project, University of Ruhuna, Wellamadama, Matara. Deadline for submission of quotations is at 2.00 P.M. on July 15, 2022
13	The Quotations shall be opened at the following address: AHEAD PROJECT, OTS-Office, University of Ruhuna, Wellamadama, Matara.
16	Other factors that will be considered for evaluation are (List and describe the methodology):
Additional Clause	
17. Payment	Payment shall be made within eight weeks after receiving the invoice and other required documents to the Office for each item supplied.
19. Liquidated Damages	The supplier should deliver the items within the delivery period stipulated in the schedule of requirements. The supplier should make payment to the purchaser for each week of delay in delivery, at the rate of 0.5% of the contract sum per each week of delay, up to a maximum of 10% of the contract sum.
21. Performance Security	If required, the supplier should forward performance security of 10% of the total contract price. This should be issued in favor the Vice Chancellor, university of Ruhuna, valid for one year period from date of acceptance of the order.

Dr. D. M. L. Virishori
 Director - Procurement
 Academic Technical Secretariat
 AHEAD Project
 University of Ruhuna
 Wellamadama, Matara.

Section III: Schedule of Requirements

Note:

1. Bidders a) May quote whole item.
2. Evaluation will be based on: a) separately.

Item No.	Brief Description of the Goods	Specification and Requirements	Quantity	Delivery Period	Bidder's offered Delivery Date
1	Electric Saw, Necropsy / Oscillating	Annexure 1	5	2 Weeks	
2	Illuminators / Viewers, X-ray	Annexure 2	5	2 Weeks	
3	Saw, Portable Jig	Annexure 3	2	2 Weeks	
4	Snellen's charts	Annexure 4	40	2 Weeks	
5	Dual Scale Classic Clinical glass Mercury thermometers	Annexure 5	5	2 Weeks	
6	Digital thermometers	Annexure 6	5	2 Weeks	

Place of Delivery: Faculty of Medicine, University of Ruhuna, Karapitiya, Galle.

Dr. D. V. L. Kristiana
 Senior Director - Procurement
 Operations Technical Secretariat
 AHEAD Project
 University Of Ruhuna
 H/00, Karapitiya, Matara.

PRICE SCHEDULE –IV

Contract: Supply, Delivery, Installation, Commissioning and Testing of Laboratory Equipment for Faculty of Medicine, University of Ruhuna.

IFB No - AHEAD/RUH/STEM3/GOODS/2022/NS-182 II

Item No.	Description	Qty.	Unit Rate. (Without VAT) (Rs. & Cents)	VAT (Rs. & Cents)	Total Amount (With VAT) (Rs. & Cents)
1	Electric Saw, Necropsy / Oscillating	5			
2	Illuminators / Viewers, X-ray	5			
3	Saw, Portable Jig	2			
4	Snellen's charts	40			
5	Dual Scale Classic Clinical glass Mercury thermometers	5			
6	Digital thermometers	5			

We agree to supply and delivery the above goods in accordance with the technical specifications for a total contract price of Rupees..... (amount in figures) (..... amount in words) within the period specified in the Invitation for Quotations.

We also confirm that the warrantee/guarantee specified shall apply to the offered goods.

Ms. V. L. KRISHANI
 Deputy Director - Procurement
 Operations Technical Secretariat
 AHEAD Project
 University Of Ruhuna
 Veterinary, Ruhuna

.....
Signature of the Bidder

.....
Name of the Bidder

.....
Date

Specifications for a Electric Saw, Necropsy / Oscillating

	Purchaser's Requirements	Priority	Bidders Offer		Remarks
			Conformity		
			Yes	No	
1	The saw shall have the ability to cut hardy materials as human skulls & Perspex sheets precisely				
2	The unit shall have a Plunge blade				
3	The unit shall have sanding pad				
4	The unit shall have sandpaper (approx.10 pc. assortment: (grits 120/180)				
5	The unit shall have an accessory adapter				
6	The unit shall have an accessory case, tool bag				
7	The unit shall have feasibility to cut bones without harming soft tissue, to remove the cranial cap, to cut & sand Perspex sheets				
8	The unit shall operate on a power supply of 230V±10% and 50Hz. If the unit requires a separate stabilized power supply, the cost of same shall be quoted separately.				
9	The unit shall have oscillating approx. 6,000 - 30,000 rpm and variable speed control dial				
10	The unit shall have an oscillating angle approx. 3.2° for faster, more aggressive cutting and sanding				
11	The unit shall have reducing vibration (approx.6 m/s ²) and lower noise (approx.74db)				
12	The unit shall have operator convenience switches: Large on/off slide switch with lock-on button for operator convenience				
13	The unit shall have clamp system and adapter to fit most competitive accessories				
14	The unit shall have section & special blades to complete with section & special blades attached arbor and wrench*				
15	The unit shall have non-metallic housing reduces heat and noise				
16	The unit shall have approx. 3 m electrical cord with fused plug with doble insulated				
17	The unit shall have electricity consumption of 3.0 A or equivalent				
18	The unit shall have following stainless Steel Saw Blades & Arbors <ul style="list-style-type: none"> • Large Section Blade with Arbor (e.g. 2-1/2" (63.5 mm)) 				

	<ul style="list-style-type: none"> • Small Section Blade with Arbor (e.g. 1-3/4" (44.5 mm)) • Spinal Column Blade with Arbor (e.g. 3-7/8" (98.5 mm)) • Round Blade with Arbor (e.g. 2-1/2" (63.5 mm)) • Arbor Wrench 				
19	The unit shall be supplied with a complete and detailed set of operation and service manuals in English				
20	The unit shall have warranty for a period not less than 36 months from the date of purchase				
21	The local representative or the manufacturer shall submit a statement of compliance (or otherwise) in respect of every clause in this specification. The bidder shall also provide printed illustrated original technical literature to substantiate conformity with the same.				
22	The local representative shall be able to sign a 5-year service contract after the guarantee period and the cost of such contract shall be quoted for separately with frequently replaceable spare parts for 5 years after the period of warranty. This should include details of routine/periodic maintenance and parts to be replaced periodically, with the price breakdown				
23	The local representative shall be able to provide a list of end users in Sri Lanka for the same product				
24	The local representative shall be able to provide relevant technical assistance with satisfactory technical service team.				
25	A recommended set of spare parts with their unit prices shall be forwarded with the offer.				
26	The unit shall be delivered within 4-6 weeks				

Make & Model	Country of origin/Manufacture
Relevant catalogue & technical information attached	Information given under "Bidder's offer" is true & correct
Bidder's Name, Signature & Date	Company seal

Specifications for a Illuminator / Viewer, X-ray

	Purchaser's Requirements	Priority	Bidders Offer		Remarks
			Conformity		
			Yes	No	
1	The unit shall have the ability to visualize developed X-ray films in teaching and examinations				
2	The unit shall have one viewing point (single unit type)				
3	The unit shall operate on a power supply of 230V±10% and 50Hz. If the unit requires a separate stabilized power supply, the cost of same shall be quoted separately.				
4	The unit shall be used by either placing it on a table or mounting on a wall..				
5	The unit shall have the output noise of <70dB or accepted low output noise exhaust fans only				
6	The unit shall have minimum working time of more than 4 hours per day				
7	The unit shall have illumination and light intensity with <ul style="list-style-type: none"> ▪ LED bulbs or "daylight" fluorescent tubes (may be 04 or more). ▪ Homogeneity of the illuminated surface: ensure that all zones are adequately illuminated. ▪ On-off switch with illuminated indicator or a shielded dimmer for variable intensities 				
8	The unit shall have view box (cabinet body) with either Sturdy aluminum frame and metal casing OR extruded PVC and injection-molded ABS				
9	The unit shall have rear panel in epoxy-coated steel				
10	The unit shall have viewing area approximately 60cm (H) x 40cm (W)				
11	The unit shall have diffusing panel (screen) made of semi-translucent white-opal cast-acrylic sheet (Plexiglas); 3mm thickness; reflectors in shaped aluminum.				
12	The unit shall have diffusing panel (screen) that could be easily removed to replace lamps – no tools required				
13	The unit shall have following features in Film & base grip <ul style="list-style-type: none"> • Self-adjusting roller-gravity film grip accommodates every film thickness • Shall not scratch or tear film • open sides facilitate viewing of oversized films • bottom film ledge accommodates small film formats 				
14	The unit shall have sliding clip with nylon wire. Transversal, cursor-adjustable for film support				
15	The unit shall have the colour preferably white				
16	The unit shall have approximate weight of 10-15Kg				
17	The unit shall have a power cord with grounded chassis with 8-foot, 3-wire line cord				

18	The unit shall be supplied with a complete and detailed set of operation and service manuals in English.				
19	The unit shall have warranty for a period not less than 60 months from the date of successful installation. Such warranty shall include servicing and preventive maintenance during the period of validity.				
20	The local representative shall be able to sign a 5-year service contract after the guarantee period and the cost of such contract shall be quoted for separately with frequently replaceable spare parts for 5 years after the period of warranty. This should include details of routine/periodic maintenance and parts to be replaced periodically, with the price breakdown				
21	The local representative shall be able to provide a list of end users in Sri Lanka for the same product				
22	The local representative shall be able to provide relevant technical assistance with satisfactory technical service team.				
23	A recommended set of spare parts with their unit prices shall be forwarded with the offer.				
24	The local representative shall submit a statement of compliance (or otherwise) in respect of every clause in this specification. The bidder shall also provide printed illustrated original technical literature to substantiate conformity with the same.				
25	The unit shall be delivered and installed within 4-6 weeks				

Make & Model	Country of Manufacture
Relevant catalogue & technical information attached	Information given under "Bidder's offer" is true & correct
Bidder's Name, Signature & Date	Company seal

Specifications for a Saw, Portable Jig

	Purchaser's Requirements	Priority	Bidders Offer		Remarks
			Conformity		
			Yes	No	
1	The saw shall have the ability to use to make curved, intricate cuts in <i>Perspex</i> (which is needed to prepare museum jars in students' teaching)				
2	The unit shall be portable handling type				
3	The unit shall operate on a power supply of 230V±10% and 50Hz. If the unit requires a separate stabilized power supply, the cost of same shall be quoted separately.				
4	The unit shall have a power cable with 3-5 m in length				
5	The unit shall h bevel angle range of 45° - 90°				
6	The unit shall have high quality & sturdy housing case with handling grip with comfortable with slip resistant surface				
7	The unit shall suitable to cut Soft Wood, <i>Perspex</i> plastic, Aluminum, Mild Steel				
8	The unit shall have Universal (U-shank) or T-shank HSS (High Speed Steel) for especially <i>Perspex</i>				
9	The unit shall have shank compatible with the blade & the jig saw which provided				
10	The unit shall have precision control mechanism where device needed to minimize blade deflection for excellent cut precision				
11	The unit shall have large die-cast aluminum footplate or similar for tool-free alteration for fast-and-easy bevel adjustments				
12	The unit shall have a motor with 7 - 8 A rating, highest in class to easily handle heavy loads				
13	The unit shall have a dial to set max speed and accelerator trigger to control operating speed (variable speed control)				
14	The unit shall have a constant response circuitry for consistent performance under load				
15	The unit shall have Multi-directional blade clamp				
16	The unit shall have bright LED lighting with on/off switch preferable for enhanced visibility of cutting area (On/off switch is essential)				
17	The unit shall have a Dust blower to keep the cut line visible and free of dust				
18	The unit shall have Aluminum gearbox with insulated				

	cover				
19	The unit shall have One-Touch blade-change system to fast blade insertion and ejection lever; no need to touch hot blade				
20	The unit shall have Four orbital-action settings to vary blade strokes				
21	The unit shall have Fully counterbalanced low-vibration mechanism for smooth operation				
22	<p>The unit shall have Assorted Jig Saw Blades (Free of charge & additional pack)</p> <ul style="list-style-type: none"> • Assorted pack - 03 types of all-purpose blades minimum for thick and thin materials specially <i>Perspex</i>, wood & metal • Needed a free of charge pack & an additional pack • Needed free of charge Plastic Shoe/ Anti-Splinter Insert/ Carrying Case 				
23	The unit shall be supplied with a complete and detailed set of operation and service manuals in English				
24	The unit shall have warranty for a period not less than 60 months from the date of purchase				
25	The local representative or the manufacturer shall submit a statement of compliance (or otherwise) in respect of every clause in this specification. The bidder shall also provide printed illustrated original technical literature to substantiate conformity with the same.				
26	The local representative shall be able to provide a list of end users in Sri Lanka for the same product				
27	The local representative shall be able to sign a 5-year service contract after the guarantee period and the cost of such contract shall be quoted for separately with frequently replaceable spare parts for 5 years after the period of warranty. This should include details of routine/periodic maintenance and parts to be replaced periodically, with the price breakdown				
28	The local representative shall be able to provide relevant technical assistance with satisfactory technical service team.				
29	A recommended set of spare parts with their unit prices shall be forwarded with the offer.				
30	The unit shall be delivered within 4-6 weeks				

Make & Model	Country of origin/Manufacture
Relevant catalogue & technical information attached	Information given under "Bidder's offer" is true & correct
Bidder's Name, Signature & Date	Company seal

Ms. D. V. L. Krishan
 Deputy Director - Procurement
 Operations Technical Secretariat
 AHEAD Project
 University Of Ruhuna
 Welimada, Malwana.





Director General
 Director General
 Director General
 Director General

Specifications for Snellen's charts

	Purchaser's Requirements	Priority	Bidders Offer		Remarks
			Conformity		
			Yes	No	
1.	The unit shall be Non-reflecting, matte finish eye charts				
2.	The unit shall be designed for vision testing at 20 foot (6 meter) distance				
3.	The unit shall have Size of the chart – approximately 22” x 11”				
4.	The unit shall have an eye occluder				
5.	The unit shall have an instructions manual				

Make & Model	Country of Manufacture
Relevant catalogue & technical information attached	Information given under “Bidder’s offer” is true & correct
Bidder’s Name, Signature & Date	Company seal

Ms. D. V. L. Krishan
 Deputy Director - Procurement
 Operations Technical Secretariat
 AHEAD Project
 University Of Ruhuna
 Wellamadana, Matara.




Ms. D. V. L. Krishan
 Deputy Director - Procurement
 Operations Technical Secretariat
 AHEAD Project
 University Of Ruhuna
 Wellamadama, Matara.

Specifications for dual scale classic clinical glass mercury thermometers

	Purchaser's Requirements	Priority	Bidders Offer		Remarks
			Conformity		
			Yes	No	
1.	The unit shall have measuring range: 35.5 °C – 42.0 °C / 95 °F - 108 °F				
2.	The unit shall be glass				
3.	The unit shall have durable case				
4.	The local representative shall submit a statement of compliance (or otherwise) in respect of every clause in this specification. The bidder shall also provide printed illustrated original technical literature to substantiate conformity with the same				

Make & Model	Country of Manufacture
Relevant catalogue & technical information attached	Information given under "Bidder's offer" is true & correct
Bidder's Name, Signature & Date	Company seal

Ms. D. V. L. Krishanu
 Deputy Director - Procurement
 Operations Technical Secretariat
 ANCAD Project
 University of Bhubaneswar
 Bhubaneswar, Odisha



Specifications for digital thermometers

	Purchaser's Requirements	Priority	Bidders Offer		Remarks
			Conformity		
			Yes	No	
1.	The unit shall measure oral, rectal and underarm temperatures				
2.	The unit shall have TONE INDICATOR beeps at peak temperature while another tone indicates when temperatures are above 37 Celsius				
3.	The unit shall have LARGE EASY TO READ digital display for easy viewing and monitoring with auto shutoff				
4.	The unit shall have durable case				
5.	The unit shall have warranty for a period not less than 12 months from the date of successful installation on full parts and labour basis. Such warranty shall include servicing and preventive maintenance during the period of validity.				
6.	The cost for spare parts shall be provided separately				
7.	The local representative shall submit a statement of compliance (or otherwise) in respect of every clause in this specification. The bidder shall also provide printed illustrated original technical literature to substantiate conformity with the same				

Make & Model	Country of Manufacture
Relevant catalogue & technical information attached	Information given under "Bidder's offer" is true & correct
Bidder's Name, Signature & Date	Company seal



Ms. D. V. L. Krishani
Deputy Director - Procurement
Operations Technical Secretariat
AHEAD Project
University Of Krbhna
Karnataka

10/10/2023
10:10:10 AM
10/10/2023